DECK PLATE THICKNESS CHART						
DASH NO.	A	B	C	ITEM 1 START PL. WT.	ITEM 2 CENTER PL. WT.	ITEM 3 END PL. WT.
	1/4"	1'-5.25"	0.25"	TBD	TBD	TBD
-1	5/16"	1'-5.31"	0.19"	TBD	TBD	TBD
-2	3/8"	1'-5.38"	0.13"	TBD	TBD	TBD

PRELIMINARY NOT FOR CONSTRUCTION. FOR ILLUSTRATION ONLY.

ITEM NO.	DESCRIPTION	QTY	MATERIAL	UNIT WT.	PART NO.	SIZE	WT.
1	START PLATFORM	1	SEE ASS'Y	see chart	D-72140		see chart
2	CTR PLATFORM	1	SEE ASS'Y	see chart	D-72141		see chart
3	END PLATFORM	1	SEE ASS'Y	see chart	D-72142		see chart
4	LOAD CELL	8	SEE ASS'Y	63.00	D-29547-2	136-75-CH; NTEP NP. 88-239-P	504
5	APPROACH COPING	2	SEE ASS'Y	91.00	B-36115-A	OPTIONALLY SUPPLIED BY EWSCO	182
6	BASE PLATE	8	ASTM A36	35.70	B-35931-D		286
7	CENTERING RING	8	A/R	1.50	B-29871-E		12
8	SHIM	24	SS	0.50	A-29869-1-D	0.06 THICK	12
9	SHIM	16	SS	0.30	A-29869-2-D	0.03 THICK	5
10	HHCS	32	SS	-		0.375-13UNC X 1.25	-
11	CLAMP	24	CS, ZN PL	-	A-29870-2-B		-

ITEM NO.	DESCRIPTION	QTY	MATERIAL	UNIT WT.	PART NO.	SIZE	WT.
12	HHCS	24	SS	-		0.50-13UNC X 1.25	-
13	MS FLAT WASHER	24	SS	-		FOR 0.500ø, ONE PER CLAMP	-
14	ANCHOR BOLT	16	ASTM A36	-	#7446	0.750 X 7.00	-
15	HHCS	4	GR. 8	-		1.25-7UNC X 6.00 FULL THREAD, ZN PL	-
16	TRI-LOCK NUT	4	GR. 2H	-		1.25-7UNC, ZN PL	-
17	SERIAL NUMBER PLATE	1	SS	-	B-34821-D		-
18	EWSCO DECAL	2	-	-	A-35412-A	8.00 X 48.00	-
19	CHECK BOLT BUMPER	4	SEE ASS'Y	10.00	B-35978-B	OPTIONALLY SUPPLIED BY EWSCO	40
20	COVER PLATE, RIGHT	4	ASTM A36		B-36465		
21	COVER PLATE, LEFT	4	ASTM A36		B-36466		
22	MANHOLE	A/R	ASTM A36	-	B-36488-1	OPTIONALLY SUPPLIED BY EWSCO	-





SECTION A-A



DETAIL A
SCALE 1:5



DETAIL B
SCALE 1:5



TYPICAL TOP VIEW OF CELL ON BASE PLATE
SCALE 1:5

ATTENTION:

SEE NOTE 1 ON INSTALLATION DRAWING # D-36118

FOR NON COMMERCIAL SCALES PROVIDE AN APPROACH OF 2 FEET MINIMUM LENGTH WHICH IS LEVEL AND IN SAME PLANE AS THE PLATFORM.

FOR COMMERCIAL SCALES CHECK WITH LOCAL WEIGHTS AND MEASURES OFFICIALS BEFORE STARTING CONSTRUCTIONS.

NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY (NIST) HANDBOOK 44 REQUIRES:

A) A STRAIGHT APPROACH OF AT LEAST ONE-HALF THE LENGTH OF THE PLATFORM.

B) NOT LESS THAN 10 FEET OF APPROACH ADJACENT TO THE PLATFORM TO BE CONSTRUCTED OF CONCRETE OR SIMILAR DURABLE MATERIAL AND BE LEVEL AND IN THE SAME PLANE AS THE PLATFORM.

PRELIMINARY NOT FOR CONSTRUCTION. FOR ILLUSTRATION ONLY.

DRAWING REFERENCES:

D-72139	SLAB FOUNDATION
D-72583	ALTERNATE FOUNDATION
D-36118-D	INSTALLATION INSTRUCTIONS
B-33069-G	4 SECTION TUBING DIAGRAM

NTEP NO, 89-185PA3
CLC = 58,000 LB CLASS IIIL

FULL SIZE PLOT D-24X36

LTR	REVISION	DATE	BY

ALL SHARP CORNERS AND EDGES TO BE BROKEN

UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES

MACHINING	FABRICATION
XX = ± .01	± .13
.XXX = ± .005	

THIS DRAWING IS THE PROPERTY OF EMERY WINSLOW SCALE COMPANY SEYMOUR, CONNECTICUT. THE INFORMATION CONTAINED HEREIN IS CONFIDENTIAL AND IS NOT TO BE USED OR DISSEMINATED TO OTHERS WITHOUT THE EXPRESS WRITTEN CONSENT OF EMERY WINSLOW SCALE COMPANY.

DRAWN	CEM
DATE	05/28/08
CHECKED	
DATE	
ACAD FILENAME:	T02275
LAYERS USED:	ALL

EMERY WINSLOW SCALE COMPANY
SEYMOUR, CT U.S.A TERRE HAUTE, IN.

MODEL 84-10070-11-TA
70' X 11', 100 TON; I-BEAM DESIGN
AG OUTLINE ARRANGEMENT

SCALE	FIRST USED ON	DRAWING NO.	REV.
1:30		D- 72138-	